Tel: 01253 206700 Fax: 01253 896648 customersupport@cmemedical.co.uk cmemedical.co.uk



# **Customer Information Bulletin**

Part Numbers	Product Description	Issued by	Hellie King
100-172S	Syringe extension set	Document No	CB15-006, Rev 02
100-172SLL	Syringe extension set	Date	21/12/15

### **Purpose of Bulletin**

CME Medical would like to inform our customers of the planned replacement of our current syringe extensions sets PN100-172S and PN100-172SLL with a new version.

The current extension set is predominantly used with our  $T34^{TM}$  and  $T60^{TM}$  ambulatory syringe pumps.

From week commencing 14/12/15 all current orders will be satisfied with the substitute administration set PN30-7100.

#### Detail

The current administration set part numbers are 100-172S or 100-172SLL and will be substituted with 30-7100

The differences between the administration sets are detailed below:

Part Number	100-172S / 100-172SLL	30-7100
Set composition	Non-DEHP PVC plasticized with	PE/PVC Infusion Lines with flexible,
	TOTM, BSE-TSE free, latex free	transparent PVC outer layer and internal
	and EVA free anti-kink tubing	Polyethylene, NON DEHP, Latex free,
		anti-kink tubing
Set length	100cm	100cm
Priming volume	0.5mL	0.8mL
Additional features	ASV/ anti reflux Male luer (patient	Rotating male luer (patient end),
	end), female luer (syringe	ASV Female luer (syringe connector),
	connector)	slide clamp

### Current set PN100-172S/100-172SLL



## Subtitute Set 30-7100



# Page 1 of 2

Registered in England and Wales

CME Medical UK Limited Kincraig Business Park Kincraig Road Blackpool FY2 OPJ Tel: 01253 206700 Fax: 01253 896648 customersupport@cmemedical.co.uk cmemedical.co.uk



#### NOTE:

The substitute set can only be used with the T34 in the lockbox with syringe sizes up to 20mL. The female luer/ASV attached to a 30mL syringe tip prevents the lockbox lid closing.

The most significant change on PN30-7100 is the male rotating luer lock that attaches to the patients access device.

Please follow the advice for correct connection of the male rotating luer to patient access device:

1. Insert Male luer firmly into access device

2. Once in position rotate one quarter turn clockwise



3. Secure in place using the rotating locking collar as shown

If you have any further questions or enquiries, please contact Customer Support on 01253 206700 or email customersupport@cmemedical.co.uk.